

**FOUR RIVERS SANITATION AUTHORITY  
ADDENDUM NO. 1  
ITB #23-206  
TWO (2) NEW LAB GRADE DISHWASHERS & INSTALLATION**

**January 31, 2023**

**AD1-1 Notice**

This Addendum No. 1, dated January 31, 2023, to the Invitation to Bid, Two (2) New Lab Grade Dishwashers & Installation, supersedes all contrary and conflicting information in the above-mentioned instructions, specifications, and contract documents which are hereby supplemented or revised in certain particulars as follows:

**AD1-2 General Information**

The Authority submits Addendum No. 1 to answer a question.

**AD1-3 Question & Answer**

**Question 1:** What are the electrical requirements for the units; the Miele basis of design appears to have a 208V plug while the Labconco models are offered in 115V and 230V.

**Answer 1:** The electrical requirements are either 208V or 230V.

Bidder shall initial this Addendum No. 1 and include it with their bid.

\_\_\_\_\_ Bidder's Initials