

**FOUR RIVERS SANITATION AUTHORITY
ADDENDUM NO. 1
RFP #24-212
One (1) New Discrete Analyzer & Installation**

May 24, 2024

AD1-1 Notice

This Addendum No. 1, dated May 24, 2024, to the Invitation to Bid: One (1) New Discrete Analyzer & Installation supersedes all contrary and conflicting information in the above-mentioned instructions, specifications, and contract documents which are hereby supplemented or revised in certain particulars as follows:

AD1-2 General Information

The Authority submits Addendum No. 1 to answer a question.

AD1-3 Question & Answer

Question: Does the discrete analyzer have to hold 90 samples and 18 reagents at the same time? If the bid requires this, then a larger Gallery Plus Aqua Master would be required. If bid specifications mean hold 90 samples and 1 reagent rack with 6 reagents at one time and allows for switching racks or holding 3 reagent racks with 18 reagent and switch racks, then the basic Gallery Aqua Master will work.

Answer: The discrete analyzer does not need to hold 90 samples and 18 reagents at the same time. These requirements can be met using different configurations or layouts; there must be one configuration that can hold 90 samples, and one configuration that can hold 18 reagents.

Bidder shall initial this Addendum No. 1 and include it with their proposal.

_____Bidder's Initials