

**FOUR RIVERS SANITATION AUTHORITY  
ADDENDUM NO. 1  
RFP #25-212  
ICP-MS & INSTALLATION**

**September 22, 2025**

**AD1-1 Notice**

This Addendum No. 1, dated September 22, 2025, to the Invitation to Bid ICP-MS & Installation supersedes all contrary and conflicting information in the above-mentioned instructions, specifications, and contract documents which are hereby supplemented or revised in certain particulars as follows:

**AD1-2 General Information**

FRSA submits Addendum No. 1 to answer a question and to strike wording.

**AD1-3 Question & Answer**

**Question 1:** On page 15-3.3 A.3, it states an optional power conditioner to be included. To ensure successful operation of the ICP-MS, monitoring/conditioning the power from the wall is extremely important and we include that in every quote/bid/order for an ICP-MS. It is not a UPS, it just conditions the power. Is that acceptable to have it in the body of the quotation and not as an option?

**Answer 1:** The power conditioner will no longer be optional. Total cost must include the ICP-MS, installation, and a power conditioner.

**AD1-4 Wording to Strike:**

From Section III, 3.3, A., 3.: "Must include an ~~optional~~ power conditioner for the ICP-MS".

Bidder shall initial this Addendum No. 1 and include it with their bid.

\_\_\_\_\_ Bidder's Initials