

**FOUR RIVERS SANITATION AUTHORITY  
ADDENDUM NO. 2  
RFP #25-212  
ICP-MS & INSTALLATION**

**September 30, 2025**

**AD2-1 Notice**

This Addendum No. 2, dated September 30, 2025, to the Invitation to Bid ICP-MS & Installation supersedes all contrary and conflicting information in the above-mentioned instructions, specifications, and contract documents which are hereby supplemented or revised in certain particulars as follows:

**AD2-2 General Information**

FRSA submits Addendum No. 2 to answer questions.

**AD2-3 Questions & Answers**

**Question 1:** Would FRSA consider removing the performance bond requirement?

**Answer 1:** The performance bond requirement is waived. The awarded vendor will not be required to provide a performance bond.

**Question 2:** What level of dilution are you requiring? Do you need 400x or 1000x?

**Answer 2:** Dilution module must perform dilutions up to 1000x

Bidder shall initial this Addendum No. 2 and include it with their bid.

\_\_\_\_\_ Bidder's Initials