

**FOUR RIVERS SANITATION AUTHORITY
ADDENDUM NO. 3
ITB #25-212
ICP-MS & INSTALLATION**

October 1, 2025

AD3-1 Notice

This Addendum No. 3, dated October 1, 2025, to the Invitation to Bid ICP-MS & Installation supersedes all contrary and conflicting information in the above-mentioned instructions, specifications, and contract documents which are hereby supplemented or revised in certain particulars as follows:

AD3-2 General Information

FRSA submits Addendum No. 3 to answer a question.

AD3-3 Questions & Answers

Question 1: Can we include in our bid an autosampler diluter that will provide auto-dilution up to 400x that is not the Elemental Scientific prepFAST 3 autosampler?

Answer 1: The bid must include an Elemental Scientific prepFAST 3 autosampler diluter or equivalent that performs dilutions up to 1000x.

Bidder shall initial this Addendum No. 3 and include it with their bid.

_____ Bidder's Initials